



Mouse anti-CD45 (LCA)

Cat. No.: BMS046 (16 ml Ready-to-use)

Instructions for use

Intended Use

This antibody is designed for the specific localization of *Leukocyte Common Antigen* (LCA, CD45) in formalin-fixed, paraffin-embedded tissue sections as well as in acetone-fixed frozen tissues sections and cytological samples. Anti-CD45 antibody is intended for in vitro diagnostic use.

Specifications

Specificity:	CD45 (LCA, Leucocyte Common Antigen)
Immunogen:	For PD7/26: humane peripheral blood cells, which are cultivated in the presence of T-cell growth factor For 2B11: from T-cell lymphoma/leukaemia isolated neoplastic cells
Clone:	PD7/26 and 2B11
Isotype:	Mouse IgG1 and Mouse IgG1
Species reactivity:	Human +, others not tested

Summary and explanation

The humane Leukocyte Common Antigen (LCA) is a family of at least five glycoproteins, which are expressed on the cell surface of the majority of humane leukocytes. This antibody cocktail consists of two monoclonal antibodies directed against different epitopes of LCA. Each of both antibodies detects 94 – 96 % of human lymphocytes and monocytes. PD7/26 also reacts with Kupffer cells.

The antibody cocktail labels lymphoid cells in tonsil, spleen, thymus and bone marrow. Additionally, mast cells stain positive. Neutrophils are stained weakly positive; plasma cells usually stain negative. Non-haematopoietic tissues do not react with this antibody cocktail. Both, B- and T-cells in non-Hodgkin lymphomas and B- and T-cell leukaemias of neoplastic tissues are labelled by this antibody cocktail. Hairy cells react positive whereas cells of granulocytic origin or cells from the erythrocytes lineage are mostly negative.

Reagent provided

Mouse monoclonal antibody in TBS with carrier protein and preservative for stabilisation in the following format:

Ready-to-use: 16 ml (Cat. No. BMS046)

Dilution of primary antibody

None

Storage and handling

The antibody should be stored at 2-8°C without further dilution.

If necessary, dilutions of the antibody should be done with a suitable antibody dilution buffer (e.g. ZUC025 from Zytomed Systems). The diluted antibody should be stored at 2-8°C after use. Stability of this working solution depends on various parameters and has to be confirmed by appropriate controls.

The antibody provided is stable until the expiry date indicated on the label, if stored at 2-8°C. Do not use product after the expiry date. Positive and negative controls should be run simultaneously with all specimens. If unexpected staining is observed which cannot be explained by variations in laboratory procedures and a problem with the antibody is suspected, contact Zytomed Systems' technical support or your local distributor.

Precautions

Use through qualified personnel only.

Wear protective clothing to avoid contact of reagents and specimens with eye, skin and mucous membranes. If reagents or specimens come in contact with sensitive area, wash with large amounts of water.

Microbial contamination of the reagent must be avoided, since otherwise non-specific staining may occur.

ProClin300 is used for stabilisation. Material safety data sheets (MSDS) are available upon request.

Staining procedure

Refer to the following table for conditions specifically recommended for this antibody. Also refer to detection system data sheets for guidance on specific staining protocols or other requirements.

Parameters	Zytomed Systems recommendations
*Pre-treatment:	<i>optional:</i> Heat Induced Epitope Retrieval (for example in citrate buffer pH 6.0 (ZUC028))
*Control tissue	Tonsil or other lymphatic tissue
*Working dilution	None
*Incubation time	30 - 60 minutes

Quality control

The recommended positive control tissue for this antibody is tonsil. We recommend carrying out a positive and a negative control with every staining run. Please refer to the instructions of the detection system for guidance on general quality control procedures.

Troubleshooting

If you observe unusual staining or other deviations from the expected results please read these instructions carefully, refer to the instructions of the detection system for relevant information or contact your local distributor.

Expected results

The antibody stains positive in the cytoplasmic membrane of lymphoid cells in formalin-fixed, paraffin-embedded tissue. Certain Hodgkin's lymphomas and true histiocytic neoplasms can also show cytoplasmic staining. The interpretation of the results is solely the responsibility of the user. Any experimental result should be confirmed by a medically established diagnostic procedure.

Limitations of the Procedure

Immunohistochemistry is a complex technique involving both histological and immunological detection methods. Tissue processing and handling prior to immunostaining, for example variations in fixation and embedding or the inherent nature of the tissue can cause inconsistent results (Nadji and Morales, 1983). Endogenous peroxidase, alkaline phosphatase or biotin may cause non-specific staining depending on the detection system used. Tissues containing Hepatitis B Surface Antigen (HBsAg) may give false positive results with HRP (horse radish peroxidase) detection systems (Omata *et al*, 1980). Inadequate counterstaining and mounting can influence the interpretation of the results.

Zytomed Systems warrants that the product will meet all requirements described from its shipping date until the expiry date is reached, if the product is stored and utilised as recommended. No additional guarantees can be given. Under no circumstances shall Zytomed System be liable for any damages arising out of the use of the reagent provided.

Performance characteristics

Zytomed Systems has conducted studies to evaluate the performance of the antibody utilising a standard detection system. The product has been found to be sensitive and specific to the antigen of interest with minimal or no cross-reactivity.

Bibliography

Warnke RA et al. N Eng J Med 309:1275 ff, 1983

Hale G et al. In McMichael AJ et al. Leucocyte Typing III.

White Cell differentiation antigens Oxford Univ Press, 811 ff, 1987

Kurtin Pj et al. Hum Pathol 16:353 ff, 1985

Parravicini CL et al. In McMichael AJ et al. Leucocyte Typing III.

White Cell differentiation antigens Oxford Univ Press, 822 ff, 1987

Nadji M and Morales AR Ann N.Y. Acad Sci 420:134-9, 1983






Omata M et al. Am J Clin Pathol 73(5): 626-32, 1980

January 29, 2014

Rev: A0114

Doc: DBE_BMS046

Explanations of the symbols on the product label:

REF	Bestellnummer Catalog Number Reference du catalogue		Verwendbar bis Use By Utiliser jusque		Gebrauchsanweisung beachten Consult Instructions for use Consulter les instructions d'utilisation
LOT	Chargenbezeichnung Batch Code Code du lot		Lagerungstemperatur Temperature Limitation Limites de température	RUO	Nur für Forschungszwecke For Research Use Only Pour la recherche uniquement
IVD	In vitro Diagnostikum In Vitro Diagnostic Medical Device Dispositif médical de diagnostic in vitro		Achtung Warning Attention		Hersteller / Manufacturer / Fabricant Zytomed Systems GmbH • Anhaltinerstraße 16 14163 Berlin, Germany • Tel: (+49) 30-804 984 990 www.zytomed-systems.com