
# Parvalbumin, RMab

Clone: EP300 Rabbit Monoclonal

C € IVD





Inset: IHC of Parvalbumin on a FFPE Kidney Tissue

# **Intended Use**

For In Vitro Diagnostic Use.

This antibody is intended for use in Immunohistochemical applications on formalin-fixed paraffin-embedded tissues (FFPE), frozen tissue sections and cell preparations. Interpretation of results should be performed by a qualified medical professional.

\*The Parvalbumin antibody, clone EP300, has been manufactured using Epitomics RabMab® technology covered under Patent No.'s 5,675,063 and 7,402,409.

# **Immunogen**

Synthetic peptide corresponding to residues of human Parvalbumin protein.

# **Summary and Explanation**

Parvalbumin is a calcium-binding albumin protein with low molecular weight (typically 9-11 kDa), structurally related to calmodulin and troponin C and a stable protein involved in calcium signaling. Parvalbumin plays a role in many physiological processes, namely cell-cycle regulation, second messenger production, muscle contraction, organization of microtubules and phototransduction. Parvalbumin is localized in fast-contracting muscles, where its levels are highest, as well as in the brain and some endocrine tissues. In normal kidney, parvalbumin has been shown to be limited to the distal tubular and collecting duct cells (the intercalated cells).

Studies have demonstrated that parvalbumin may be a suitable IHC marker for distinguishing primary and metastatic chromophobe carcinoma from conventional (clear cell) and papillary renal cell carcinoma. Parvalbumin is strongly expressed in almost all primary, as well as metastatic, chromophobe renal cell carcinoma (RCC) (100%) and oncocytoma (69%), but is essentially negative in other types of RCCs, such as clear cell RCC and papillary RCC.

Additionally, alterations in the function of parvalbumin-expressing neurons have been implicated in various areas of clinical interest such as Alzheimer's disease and age-related cognitive defects.

Antibody Type	Rabbit Monoclonal	Clone EP300		
Isotype	lgG	Reactivity Paraffin, Frozen		
Localization	Cytoplasmic	Control	Brain, Kidney, Tonsil, Lymph Node, Chromophobe RCC	
Species Reactivity H		Human		

# **Presentation**

Parvalbumin is a rabbit monoclonal antibody derived from cell culture supernatant that is concentrated, dialyzed, filter sterilized and diluted in buffer pH 7.5, containing BSA and sodium azide as a preservative.

# **Presentations**

Catalog Num.	Antibody Type	Dilution	Volume/Qty
BSB 3406	Tinto Prediluted	Ready-to-Use	3.0 mL
BSB 3407	Tinto Prediluted	Ready-to-Use	7.0 mL
BSB 3408	Tinto Prediluted	Ready-to-Use	15.0 mL
BSB 3409	Concentrated	1:25 - 1:100	0.1 mL
BSB 3410	Concentrated	1:25 - 1:100	0.5 mL
BSB 3411	Concentrated	1:25 - 1:100	1.0 mL
BSB 3412	Control Slides	Not Applicable	5 slides

# **Precautions**

- **1.** For professional users only. Ensure results are interpreted by a medical professional.
- **2.** This product contains sodium azide (NaN3), a toxic chemical which may react with plumbing to form highly explosive build-ups of metal azides. Upon disposal, flush with large volumes of water to prevent sodium azide build-up.
- **3.** Ensure proper handling procedures are used with reagent. Always wear proper laboratory equipment such as laboratory coat and gloves when handling reagents.
- **4.** Unused solution should be disposed of according to local and federal regulations.
- **5.** Do not ingest reagent. If reagent ingested, contact a poison control center immediately.
- **6.** For complete recommendations for handling biological specimens please refer to the CDC document, "Guidelines for Safe Work Practices in Human and Animal Medical Diagnostic Laboratories" (5).

### Storage

**Store at 2-8** °C. Do not use after expiration date listed on package label. Temperature fluctuations should be avoided. Store appropriately when not in use, and avoid prolonged exposure to room temperature conditions.

# **Specimen Preparation**

Paraffin sections: The antibody can be used on formalin-fixed paraffin-embedded (FFPE) tissue sections. Ensure tissue undergoes appropriate fixation to ensure best results. Pre-treatment of tissues with heat-induced epitope retrieval (HIER) is recommended using Bio SB ImmunoDNA Retriever with Citrate (BSB 0020-BSB 0023), ImmunoDNA Retriever with EDTA (BSB 0030-BSB 0033) or ImmunoDNA Digestor (BSB 0108-0112). See reverse side for complete protocol. Tissue should remain hydrated via use of Bio SB Immuno/DNA Washer solutions (BSB 0029 & BSB 0042).

**Frozen sections and cell preparations:** The antibody can be used for labeling acetone-fixed frozen sections and acetone-fixed cell preparations.

# **Staining Procedure**

- Cut and mount 3-5 micron formalin-fixed paraffin-embedded tissues on positive charged slides such as Bio SB Hydrophilic Plus Slides (BSB 7028).
- 2. Air dry for 2 hours at 58° C.
- 3. Deparaffinize, dehydrate and rehydrate tissues.
- Subject tissues to heat epitope retrieval using a suitable retrieval solution such as ImmunoDNA Retriever with Citrate (BSB 0020-BSB 0023) or EDTA (BSB 0030-BSB 0033).
- 5. Any of three heating methods may be used:

### a. TintoRetriever Pressure Cooker or Equivalent

Place tissues/slides in a staining dish or coplin jar containing the ImmunoDNA Retriever with Citrate or EDTA, and place in the pressure cooker. Add 1-2 inches of distilled water to the pressure cooker and turn heat to high. Incubate for 15 minutes. Open and immediately transfer slides to room temperature.

#### b. TintoRetriever PT Module or Water Bath Method

Place tissues/slides in a pre-warmed staining dish or coplin jar containing the ImmunoDNA Retriever with Citrate or EDTA at 95°-99° C. Incubate for 30-60 minutes.

#### c. Conventional Steamer Method

Place tissues/slides in a pre-warmed staining dish or coplin jar containing the ImmunoDNA Retriever with Citrate or EDTA in a Steamer, cover and steam for 30-60 minutes.

- After heat treatment, transfer slides in ImmunoDNA Retriever with Citrate or EDTA to room temperature and let stand for 15-20 minutes.
- 7. For manual staining, perform antibody incubation at ambient temperature. For automated staining methods, perform antibody incubation according to instrument manufacturer's instructions.
- 8. Wash slides with IHC wash buffer or DI water.
- 9. Continue IHC staining protocol.

### **Recommended IHC Protocol**

Step	ImmunoDetector AP/HRP	PolyDetector AP/HRP	PolyDetector Plus HRP	
Peroxidase/AP Blocker	5 min.	5 min.	5 min	
Primary Antibody	30-60 min.	30-60 min.	30-60 min.	
1st Step Detection	10 min.	30-45 min.	15 min.	
2nd Step Detection	10 min.	Not Applicable	15 min.	
Substrate-Chromogen	5-10 min.	5-10 min.	5-10 min.	
Counterstain	Varies	Varies	Varies	

### **Product Limitations**

Due to inherent variability present in immunohistochemical procedures (including fixation time of tissues, dilution factor of antibody, retrieval method utilized and incubation time), optimal performance should be established through the use of positive and negative controls. Results should be interpreted by a medical professional.

## References

- 1. Cates MS,et al. Metal-ion affinity and specificity in EF-hand proteins: coordination geometry and domain plasticity in parvalbumin. Structure.1999; 7 (10): 1269–78
- 2. Cates MS, eta al. Molecular mechanisms of calcium and magnesium binding to parvalbumin. Biophys. J. 2002; 82 (3): 1133–46.
- 3. Martignoni G, et al. Parvalbumin is constantly expressed in chromophobe renal carcinoma. Mod Pathol. 2001 Aug;14(8):760-7.
- 4. Verret L, et al. Inhibitory interneuron deficit links altered network activity and cognitive dysfunction in Alzheimer model". Cell. 2012; 149 (3): 708–21.
- 5. U.S. Department of Health and Human Services: Centers for Disease Control and Prevention. Guidelines for Safe Work Practices in Human and Animal Medical Diagnostic Laboratories. Supplement / Vol. 61, January 6, 2012.

# Symbol Key / Légende des symboles/Erläuterung der Symbole

EMERGO EUROPE Prinsessegracht 20 2514 AP The Hague The Netherlands	N _8°C	Storage Temperature Limites de température Zulässiger Temperaturbereich	1	Manufacturer Fabricant Hersteller	REF	Catalog Number Référence du catalogue Bestellnummer
In Vitro Diagnostic Medical Device Dispositif médical de diagnostic in vitro In-Vitro-Diagnostikum	i	Read Instructions for Use Consulter les instructions d'utilisation Gebrauchsanweisung beachten		Expiration Date Utiliser jusque Verwendbar bis	LOT	Lot Number Code du lot Chargenbezeichnung



