

Chromogranin A

Catalog # 10117.05, .10 or .50 (0.5ml or 1.0ml concentrated, or 5 ml RTU)

INTENDED USE

- **In Vitro:** This product is intended for qualitative immunohistochemistry with normal and neoplastic, formalin-fixed, paraffin embedded tissue sections, to be viewed by light microscopy.
- **Description:** Chromogranin A is present in a wide variety of endocrine tissues including the pituitary, pancreas, hypothalamus, thymus, thyroid, intestine and parathyroid. Co-expression of chromogranin and NSE is typical of neuroendocrine neoplasms. Most pituitary adenomas and prolactinomas readily express chromogranin.
- **Expected Staining Pattern:** Cytoplasmic
- **Positive Control:** Pancreas

MATERIALS PROVIDED

- **Chromogranin A specific monoclonal antibody:** #10117.10, (or .05) is tissue culture supernatant, concentrated, with 0.09% sodium azide.
- **Host:** Mouse
- **Epitope:** Not determined.
- **Species Reactivity:** Human.
- **Clone:** LK2H10
- **Immunogen:** Not determined.
- **Sterility:** This product is not sterile.

MATERIALS REQUIRED, BUT NOT PROVIDED

- Detection system
- Chromogen/substrate system

METHODS AND PROCEDURES

- **Tissue Section Pretreatment:** Staining of formalin-fixed tissue sections requires treating in boiling 10mM citrate buffer, pH 6.0 for 10-20 minutes.
- **Recommended Pretreatment Kit:** HISTOLS®-Pre-Treatment IHC Kit (cat# 30016), containing the reagents below:
HISTOLS®-Citrate Buffer (cat#30010)
HISTOLS®-Peroxidase Blocking Solution (cat# 30013)
HISTOLS®-BBPS (Background Blocking Protein Solution)(cat# 30012)
- **Dilution of Concentrated Antibody:** 1:100 in antibody diluent.
- **Primary Antibody Incubation Time:** 60 minutes at RT.
- **Recommended Detection System:** HISTOLS®-MR (anti-mouse&rabbit Ig) (cat# 30011)
- **Visualization:** To detect antibody binding sites, follow the instructions provided with the chromogen/substrate system.

STORAGE AND STABILITY

This product contains sodium azide and is stable when stored at 2-8°C. Do not use after expiration date indicated on label of the product. If reagent is not stored as recommended, performance must be validated by the user.

For research use only!

