


CD44 Control Slides





Intended Use

For In Vitro Diagnostic Use.

Summary and Explanation

The CD44 protein is a cell-surface glycoprotein involved in cell-cell interactions, cell adhesion and migration. CD44 is also known as Homing-cell adhesion molecule and Phagocytic glycoprotein-1. A specialized sialofucosylated glycoform of CD44 called HCELL is found natively on human hematopoietic stem cells and functions as a "bone-homing receptor", directing migration of human hematopoietic stem cells and mesenchymal stem cells to bone marrow.

This protein participates in a wide variety of cellular functions including lymphocyte activation, recirculation and homing, hematopoiesis, and tumor metastasis. Transcripts for this gene undergo complex alternative splicing that results in many functionally distinct isoforms; however, the full-length nature of some of these variants have not been determined. Splice variants of CD44 on Colon Cancer cells display the HCELL glycoform, which mediates binding to vascular E-selectin under hemodynamic flow conditions, a critical step in Colon Cancer metastasis. In addition, variations in CD44 are reported as cell surface markers for some breast and prostate cancer stem cells and have been seen as an indicator of increased survival time in Epithelial Ovarian Cancer patients.

Presentation

Five slides of CD44 positive tissues, each mounted on Hydrophilic Plus Slides, provided in a plastic mailer.

Catalog No.	Quantity		
BSB-9094-CS	5 slides		
BSB 6239	5 slides		

Storage Store at 20-25°C

Precautions

- 1. For professional users only. Results should be interpreted by a qualified medical professional.
- 2. Ensure proper handling procedures are used with this reagent.
- 3. Always wear personal protective equipment such as a laboratory coat, goggles, and gloves when handling reagents.
- 4. Dispose of unused solution with copious amounts of water.
- 5. Follow safety precautions of the heating device used for epitope retrieval (TintoRetriever Pressure Cooker or similar).
- 8. For additional safety information, refer to Safety Data Sheet for this product.
- 9. For complete recommendations for handling biological specimens, please refer to the CDC document, "Guidelines for Safe Work Practices in Human and Animal Medical Diagnostic Laboratories" (see References in this document).

Stability

This product is stable up to the expiration date on the product label. Do not use after expiration date listed on package label.

IHC Protocol

- 1. Subject tissues to heat induced epitope retrieval (HIER) using a suitable retrieval solution such as ImmunoDNA Retriever with Citrate (BSB 0020-BSB 0023) or EDTA (BSB 0030-BSB 0033).
- 2. Any of three heating methods may be used:

a. TintoRetriever Pressure Cooker or Equivalent

Place tissues/slides in a staining dish or coplin jar containing the ImmunoDNA Retriever with Citrate or EDTA and place on trivet in the pressure cooker. Add 1-2 inches of distilled water to the pressure cooker and turn heat to high. Incubate for 15 minutes. Open and immediately transfer slides to room temperature.

b. TintoRetriever PT Module or Water Bath Method

Place tissues/slides in a pre-warmed staining dish or coplin jar containing the ImmunoDNA Retriever with Citrate or EDTA at 95°-99° C. Incubate for 30-60 minutes.

c. Conventional Steamer Method

Place tissues/slides in a pre-warmed staining dish or coplin jar containing the ImmunoDNA Retriever with Citrate or EDTA in a steamer, cover and steam for 30-60 minutes.

- 3. After heat treatment, transfer slides in ImmunoDNA Retriever with Citrate or EDTA to room temperature and let stand for 15-20 minutes.
- 4. For manual staining, perform antibody incubation at ambient temperature. For automated staining methods, perform antibody incubation according to instrument manufacturer's instructions.
- 5. Wash slides with ImmunoDNA washer or DI water.
- 6. Continue IHC staining protocol. Wash slides between each step with ImmunoDNA washer solution.

Abbreviated Immunohistochemical Protocol

Abbreviated illilianonistochemicat i rotocot								
Step	ImmunoDetector AP/HRP	PolyDetector AP/HRP	PolyDetector Plus HRP					
Peroxidase/AP Blocker	5 min.	5 min.	5 min					
Primary Antibody	30-60 min.	30-60 min.	30-60 min.					
1st Step Detection	10 min.	30-45 min.	15 min.					
2nd Step Detection	10 min.	Not Applicable	15 min.					
Substrate- Chromogen	5-10 min.	5-10 min.	5-10 min.					
Counterstain / Coverslin	Varies	Varies	Varies					

Abbreviated IF Protocol

Step	Incubation Time		
Rinse slides in IF wash buffer	5 minutes		
Drain and wipe excess IF wash buffer off slide			
Conduct remaining steps in the dark			
Apply Antibody	30-60 minutes		
Rinse with 3 changes of IF wash buffer	3x15 minutes each		
Coverslip with IF mounting medium			

Mounting Protocols

For detailed instructions using biodegradable permanent mounting media such as XyGreen PermaMounter (BSB 0169-0174) or organic solvent based resin such as PermaMounter (BSB 0094-0097), refer to PI0174 or PI0097.

Product Limitations

Due to inherent variability present in immunohistochemical procedures (including fixation time of tissues, dilution factor of antibody, retrieval method utilized, and incubation time), optimal performance should be established through the use of positive and negative controls. Results should be interpreted by a qualified medical professional.

References

- 1. Li et al. Cell Research. 2007;17:3-14
- 2. Sillanpää S, et al. Clin Cancer Res. 2003;9(14):5318-24
- 3. Yasuda M, et al. Histol. Histopathol. 2003;17(3): 945-50
- 4. Ponta H, et al. Nat. Rev. Mol. Cell Biol. 2003;4(1): 33-45
- 5. U.S. Department of Health and Human Services: Centers for Disease Control and Prevention. Guidelines for Safe Work Practices in Human and Animal Medical Diagnostic Laboratories. Supplement / Vol. 61, January 6, 2012.

https://www.cdc.gov/mmwr/pdf/other/su6101.pdf

Symbol Key / Légende des symboles/Erläuterung der Symbole

EC REP	QAdvis EAR AB Ideon Science Park Scheelevägen 17 SE-223 70 Lund, Sweden	\	Storage Temperature Limites de température Zulässiger Temperaturbereich		Manufacturer Fabricant Hersteller	REF	Catalog Number Référence du catalogue Bestellnummer
IVD	In Vitro Diagnostic Medical Device Dispositif médical de diagnostic in vitro In-Vitro-Diagnostikum	(i	Read Instructions for Use Consulter les instructions d'utilisation Gebrauchsanweisung beachten	\subseteq	Expiration Date Utiliser jusque Verwendbar bis	LOT	Lot Number Code du lot Chargenbezeichnung

